



Quality Club Total IgE General Monthly Report


Test Period: August 2012

Results from all labs

N (except outliers)	171
Median	318 kU/l
Mean	315 kU/l
SD	25,9 kU/l
CV	8,23 %
Mean - 2 SD	263 kU/l
Mean + 2 SD	367 kU/l

Sample A (Lot No. AX1B6)

No of participating labs: 177



Results from all labs

N (except outliers)	159
Median	104 kU/l
Mean	104 kU/l
SD	7,29 kU/l
CV	7,02 %
Mean - 2 SD	89,2 kU/l
Mean + 2 SD	118 kU/l

Sample B (Lot No. AX5B6)

No of participating labs: 171

